

NEWS FROM
Loop AND Learn

Master Your DIY Closed-Loop Insulin Delivery System



27 August 2023

[SUBSCRIBE](#)

[BACK ISSUES](#)

[View larger print version](#)

In This Issue

- Upcoming Open Mic
- iAPS Videos
- Flagging Non-Pump Insulin
- Report Device Issues
- Q&A From Our Facebook Group
- Keep Your Loop Looping



OPEN MIC



iAPS presentations are hosted as a service to the iAPS community and questions about iAPS should be directed to the **iAPS** group.

Did you watch **the last two** iAPS Open Mics and start using iAPS? Still thinking about trying iAPS? Curious about how to deal with common scenarios? Please join us for another informative session as Magnus Reintz and Theresa Hastings answer your iAPS questions.

PAGE 1/2

Flagging Non-Pump Insulin


Occasionally, we want to look back and find particular instances of non-pump insulin we've entered. These can be flagged by adding .001 to the entry. An example of its utility is when Loopers use fake bolus entries to reduce negative IOB. *Please note that negative IOB can be an indicator that settings need adjustment.

Q & A from our Facebook Group

Calling attention to a helpful post

Is there a way for parents to get alerted if Loop is not Looping?

[See what our experts said!](#)



We used to have to enter "fake boluses" through Apple Health.

Now we can do it directly in the Loop app!



Device Issues?

We encourage everyone to [report device issues to the FDA](#). This is a very wise thing to do, so when there are significant numbers of reports, serious problems can be flagged.

Keep Your Loop Looping

Here are some easy things you can do to keep your Loop up and running:

- Use the newest version of Loop: [Version 3.2.2](#)
- Be sure you've signed the updated (as of June) [Apple Developer Agreement](#)